

Proficiency test through interlaboratory comparison
Radiated emission measurements in the 30 MHz to 6000 MHz frequency range
Scheme of the proficiency test PTC(RE-30-6000-II)

Carlo Carobbi (Università degli Studi di Firenze)
Michele Borsero (Istituto Nazionale di Ricerca Metrologica, Torino)
Giuseppe Vizio (Istituto Nazionale di Ricerca Metrologica, Torino)
Alessio Bonci (Università degli Studi di Firenze)
Marco Cati (Powersoft S.p.A., Firenze)
Carlo Panconi (Elettroingegneria, Pistoia)

Rev. 2 – November 18, 2015

1. Scope

1.a) This document describes the participation scheme to a proficiency test performed through an interlaboratory comparison of radiated emission measurements. The scheme includes:

- The description of the interlaboratory comparison;
- The selection criteria of the participants and the terms of admission to the proficiency test;
- The description of the technique adopted for the statistical analysis of the results of the interlaboratory comparison;
- The instructions to the participating laboratory (briefly, Laboratory) on how to perform measurements;
- The description of the method by which the results of the proficiency test are registered by the Laboratory and by the Coordinator of the proficiency test;
- The test reports issued by the Laboratory and the Coordinator;
- The registration form.

1.b) The last revision of the present document can be downloaded from the following URL:
<http://www.emc.unifi.it/CMpro-v-p-26.html>

2. Coordinator

2.a) The Coordinator of the proficiency test is Carlo Carobbi, from Università degli Studi di Firenze. The Coordinator relies on the technical and scientific support from:

- Michele Borsero, Istituto Nazionale di Ricerca Metrologica (Torino);
- Giuseppe Vizio, Istituto Nazionale di Ricerca Metrologica (Torino);
- Alessio Bonci, (Università degli Studi di Firenze);
- Marco Cati, Powersoft S.p.A. (Firenze);
- Carlo Panconi, Elettroingegneria (Pistoia).

2.b) The contact details of the Coordinator are reported below:

Carlo Carobbi

Dipartimento di Ingegneria dell'Informazione

Università degli Studi di Firenze

Via S. Marta, 3 – 50139 Firenze

Phone: +39 055 2758501

Mob. phone: +39 329 6509116

e-mail: carlo.carobbi@unifi.it

3. Type of interlaboratory comparison

3.a) The interlaboratory comparison consists in the comparison of the measurements of a travelling standard (Sample) provided by the Coordinator. Each Laboratory makes a quantitative examination (Measurement) of the Sample thus providing the Coordinator with a measurement result.

3.b) The Coordinator designed and assembled the Sample.

3.c) The Coordinator assigned to the Sample two reference values and the corresponding uncertainties. One reference value, X_{cal} (and its expanded uncertainty U_{cal}), was obtained by the Coordinator through measurements and numerical predictions. The reference value X_{cal} (and U_{cal}) is known before the start of the proficiency test. The other reference value, x^* (and its expanded uncertainty U^*), is obtained by the Coordinator through the statistical analysis of the measurement results provided by the Laboratories during the proficiency test. The reference value x^* (and U^*) will be known at the end of the proficiency test, after that the last participating Laboratory has sent its measurement results.

3.d) The scheme of participation in the proficiency test is sequential and it is illustrated in Fig. 1. The Coordinator passes the Sample to the 1st participating Laboratory. The 1st Laboratory takes the measurement thus obtaining the 1st measurement result. Then, the 1st Laboratory passes the Sample to the 2nd Laboratory which, in turns, makes the measurement and determines the 2nd measurement result. The 2nd Laboratory passes the Sample to the 3rd Laboratory which determines the 3rd measurement result, and so on. The last Laboratory passes back the Sample to the Coordinator. The proficiency test is completed when the last participating Laboratory has sent its measurement results to the Coordinator.

3.e) The measurement result provided by each Laboratory consists of a measured value x and its expanded uncertainty U_{lab} . The measurement result provided by each Laboratory shall be compared against one or both the reference values assigned by the Coordinator (see below).

3.f) Two options are available to Laboratories:

- **Option I:** The transmission of the test report from the Coordinator to the Laboratory will take place only after that the proficiency test is concluded. No communication of the results of the proficiency test shall be done by the Coordinator to the Laboratory in the time period between the beginning and the conclusion of the proficiency test. The measurement result provided by each Laboratory shall be compared against the reference value x^* (and its expanded uncertainty U^*).

- **Option II:** The Laboratory sends the measurement result to the Coordinator and the Coordinator immediately communicates to the Laboratory the outcome of the comparison between the measurement result and the reference value X_{cal} and its expanded uncertainty U_{cal} . The communication of the outcome of the comparison between the measurement result provided by the Laboratory, x and U_{lab} , with x^* and its expanded uncertainty U^* will be made at the conclusion of the proficiency test. The Coordinator will therefore send two test reports:
 - The first test report with the outcome of the comparison between the measurement result, x and U_{lab} , and the reference value X_{cal} and U_{cal} .
 - The second test report with the outcome of the comparison between the measurement result, x and U_{lab} , and the reference value x^* and U^* .

3.g) A higher participation fee is requested to the Laboratory willing to choose Option II, in addition to the agreement to not disclose the reference value X_{cal} and U_{cal} assigned by the Coordinator (see the Contract in Annex A – Italian – or Annex B – English).




Fig. 1: Sequence by which the Sample is passed from the Coordinator to the Laboratories and from the Laboratories to the Coordinator.

3.h) The Laboratory has one (1) week available to perform the measurement and one (1) week to communicate the measurement result to the Coordinator. Late results will not be accepted nor processed by the Coordinator.

3.i) In case that a Laboratory is willing to submit more than one set of measurement results (e.g. because the Laboratory wants to assess the performance of different instrumentations) then the Laboratory shall contact the Coordinator in order to determine the appropriate scheduling and participation fee. In such case more than one test report shall be issued by the Laboratory, one for each submitted set of measurement results. Any request for more than one test report must be individually evaluated by the Coordinator (the Laboratory shall contact the Coordinator to this purpose) and it will imply, if viable, a higher cost (250 Euro for any additional test report).

3.j) It is intended that the subscription of the contract in Annex A (Italian) or Annex B (English) allows for the submission of a single test report, both by the Laboratory and by the Coordinator.

4. Admission requirements

4.a) The present scheme applies to Electromagnetic Compatibility (EMC) test Laboratories that:

- Can make radiated emission measurements in accordance to the methods described in the following clauses of the standard EN 55016-2-3:2010:
 - o §7.3 (30-1000 MHz, semi-anechoic room or open area test site, at 3 or 10 m distance) or in §7.4 (30-1000 MHz, fully-anechoic room or absorber-lined semi-anechoic room or absorber-lined open area test site, at 3 m distance) and,
 - o §7.6 (1-18 GHz, fully-anechoic room or absorber-lined semi-anechoic room or absorber-lined open area test site, at 3 m distance)in the frequency range between 30 and 6000 MHz;

- Have evaluated the measurement uncertainty U_{lab} of the measurement method mentioned in the previous clause - and U_{lab} is less than or equal to the corresponding U_{cispv} value reported in EN 55016-4-2:2011, i.e. $U_{lab} \leq 6,3$ dB in the frequency range between 30 and 1000 MHz in a semi-anechoic room or open area test site, $U_{lab} \leq 5,3$ dB in the frequency range between 30 and 1000 MHz in a fully-anechoic room or absorber-lined semi-anechoic room or absorber-lined open area test site, $U_{lab} \leq 5,2$ dB in the frequency range between 1 and 6 GHz in a fully-anechoic room or absorber-lined semi-anechoic room or absorber-lined open area test site.

4.b) Accreditation to ISO/IEC 17025 is not required for admission to the proficiency test. The Coordinator designed the present scheme assuming participation of both accredited and non-accredited Laboratories.

4.c) The Coordinator starts the proficiency test if there are at least five participating Laboratories. The maximum number of participating laboratories is twenty five (25) which corresponds to a total duration of the proficiency test of less than one year.

4.d) The Laboratory that is willing to participate in the proficiency test shall:

- Fill, print, sign, scan and send by certified e-mail to the certified e-mail address dinfo@pec.unifi.it the contract in Annex A (Italian) or Annex B (English).
- Designate a Technical Responsible and, if possible, his/her Deputy by filling and sending by e-mail to the Coordinator (carlo.carobbi@unifi.it) the form in Annex C. The Technical Responsible shall sign the test report submitted by the Laboratory to the Coordinator, in addition he/she will be the reference person for correspondence with the Coordinator. The Laboratory shall communicate to the Coordinator its address through the same Annex C.
- Select the period during which the measurement will be performed by using the Doodle link <http://doodle.com/6fy7265i6z9e4i4c>. Use the Laboratory name when making the selection, do not use the code assigned by the Coordinator.

4.e) The Laboratory shall observe the following shipping rules:

- Shipping of the Sample from the Coordinator to the Laboratory is in charge of the Coordinator;
- Shipping of the Sample from the Laboratory to the next Laboratory or to the Coordinator is in charge of the Laboratory;

- Shipments shall be done by means of an express courier;
- The same packaging used by the Coordinator shall be used by the Laboratory.

Information to each Laboratory about the address and contact details of the previous and the next Laboratories in the round is provided by the same Doodle link. Be accurate when inserting the address and contact details in the “Comment” field provided by Doodle.

4.f) Handle with care the travelling Sample. A damage to the Sample will cause a delay and eventually the interruption of the proficiency test. Each Laboratory shall verify by inspection the mechanical integrity of the Sample. Possible defects or damages, proven or suspected, shall be immediately notified to the Coordinator. A verification of the electrical performance of the Sample is also envisaged (see §7).

4.g) The Coordinator assigns a code to the Laboratory. The same code will be used to identify the Laboratory in correspondence and in the test reports. The code is as follows:

PTC(RE-30-6000-II)LAB(#)

The code is the combination of a general part – PTC(RE-30-6000-II) – that identifies the measurement method, and therefore a homogenous set of measurement results, and a specific part – LAB(#) – that identifies a particular Laboratory.

5. Statistical analysis of the measurement results

5.a) The statistical analysis is based on the zeta-scores (symbol ζ) performance statistics (see §7.7 of ISO 13528:2005). The measurement result x_i , in dB(μ V), provided by the i -th Laboratory ($i = 1, 2, \dots, p$, where p is the number of participating Laboratories) is compared with the value X , in dB(μ V/m), assigned by the Coordinator. The standard uncertainty of x_i is $u_{xi} = (U_{lab})_i / 2$ where $(U_{lab})_i$, in dB, is the expanded uncertainty stated by the i -th Laboratory (see §3). The standard uncertainty of X is $u_X = U / 2$, where U , in dB, is the expanded uncertainty obtained multiplying the standard uncertainty by a coverage factor $k = 2$ (which corresponds to a coverage probability of about 95 %, assuming a normal distribution) that the Coordinator assigned to the reference value X . The Coordinator calculates the following measure ζ_i of relative deviation between x_i and X :

$$\zeta_i = \frac{x_i - X}{\sqrt{u_{xi}^2 + u_X^2}}. \quad (1)$$

The value of ζ_i is calculated for each Laboratory and for each investigated frequency. Therefore as many values of ζ_i will be calculated as the number of investigated frequencies (ten frequencies investigated, ten values of ζ_i for the i -th Laboratory). The measurement result provided by the i -th Laboratory will produce a warning signal if, at least at one frequency, we have ζ_i less than -2 or greater than $+2$. The measurement result provided by the i -th Laboratory will produce an action signal if, at least at one frequency, we have ζ_i less than -3 or greater than $+3$. If at all frequencies we have ζ_i greater than -2 and less than $+2$ then the measurement result provided by the i -th Laboratory will not give evidence of any anomaly.

5.b) The reference values X and U that the Coordinator uses to evaluate the performance of a Laboratory may be only x^* and U^* or both x^* and U^* and X_{cal} and U_{cal} , depending on the Option I or II chosen by the Laboratory (see §3).

5.c) The standard uncertainty of the reference value $X = x^*$ is $u_x = U^*/2 = 1,25 \cdot s^*/\sqrt{p}$. Both x^* and s^* are obtained from the statistical analysis of the measurement results provided by the p laboratories.

5.d) The values of x^* and s^* are obtained by the Coordinator by using the robust analysis (Algorithm A) described in Annex C of ISO 13528:2005. The robust analysis is based on an iterative calculation. At the first step of iteration

$$x^* = \text{median of } x_i \quad (i = 1, 2, \dots, p) \quad (2)$$

and

$$s^* = 1,483 \cdot \left\{ \text{median of } |x_i - x^*| \right\} \quad (i = 1, 2, \dots, p). \quad (3)$$

NOTE 1: The factor 1,483 which appears in (3) represents the ratio between the standard deviation σ and the median of the absolute deviations from the median, MAD , assuming normal distribution. It is indeed possible to show that in the case of symmetric distribution, $MAD/\sigma = \Phi^{-1}(3/4)$, where Φ is the cumulative distribution function. In the case of normal distribution $\Phi^{-1}(3/4) = 0,6745$ and therefore $\sigma = 1,4826 \cdot MAD$.

NOTE 2: The factor 1,25 that appears in the formula $U^*/2 = 1,25 \cdot s^*/\sqrt{p}$ represents the ratio between the standard deviation of the median and the standard deviation of the mean (see the note in §5.5.2 of ISO 13528:2005). Therefore x^* and s^* are interpreted as the median and the standard deviation of the mean of the measurement results, respectively.

NOTE 3: The Coordinator compares the reference value assigned to the electromagnetic field generated by the Sample with the average of the measurement results provided by the Laboratories by using the performance statistic described in §5.7 of ISO 13528:2005 where x^* and s^* are obtained by using the robust analysis (Algorithm A) described in Annex C of the standard ISO 13528:2005. In this way the Coordinator can highlight the presence of a possible bias affecting the reference value (due to an uncorrected systematic effect) or affecting the measurement results produced by the Laboratories (because inherent in the standard measurement method).

6. Characteristics of the Sample

6.a) The Sample is an electromagnetic field source made of the combination of a battery operated comb generator and an antenna. The comb generator and the battery are embedded in the antenna.



(a)



(b)

Fig. 1: (a) – Picture of the Sample; (b) – Picture of the Sample inside its case. Note that also the power supply of the Sample is included in the case (two items: 1) Sample, 2) Power supply).

6.b) The Sample is equipped with a button through which the Laboratory can turn on and turn off the comb generator. In order to turn on/off the comb generator press the button for at least five seconds. The (green) led in the button flashes when the comb generator is on.

6.c) Three hours of continuous and reliable operation of the comb generator are permitted by the fully charged battery. The comb generator automatically turns off when the charge level drops below the limit for reliable operation.

6.d) The Sample is equipped with a USB connector for battery recharge. The power supply is provided by the Coordinator. Three hours are needed to fully recharge the battery. During battery recharge the led intermittently flashes. If the led is continuously on then the battery is fully charged.

6.e) The comb generator does not require warm up prior to measurement.

6.f) The harmonics' spacing is 50 MHz and the first available harmonic in the frequency range from 30 MHz to 6000 MHz is at 50 MHz.

6.g) The Coordinator identifies the harmonics to be measured through their ordinal number and the approximate frequency value. For example: the 5th harmonic at approximately 250 MHz.

6.h) What identifies the harmonic is its order not its frequency. Frequencies are given only for guidance.

6.i) The combination of the lock is 1839. Input the figures from left to right when looking the lock from the side of the brand "THEMA".

7. Measurement procedure

7.a) Radiated electromagnetic field measurement must be preceded by a preliminary verification of the correct operation of the Sample by using the magnetic field probe provided by the Coordinator, a short section of coaxial cable (length less than 1 m, not provided by the coordinator) and a receiver (spectrum analyzer or EMI receiver). The verification shall be as follows:

- Connect the probe to the input of the receiver through the short section of coaxial cable.
- Put the Sample on the same table used for radiated emission testing of table-top equipment.
- Turn on the Sample. The Sample shall be fed by its internal battery (the Sample shall not be connected with the power supply).
- Place the probe in the position sketched in Fig. 2. In particular the probe shall be positioned so that its cable is perpendicular to the axis of the Sample and at half height of the Sample. The plane of the loop shall be the one containing the probe cable and the axis of the Sample. The probe shall touch the plastic wall of the Sample.
- Use your hand to support the probe. No special care is required.




Fig. 2: Verification of the Sample by using a magnetic field probe (the one provided by the Coordinator).

- Measure the power P_m that the probe delivers to the input of the receiver at the frequency of the 20th harmonic (about 1000 MHz). Register the values of P_m (in dBm, rounded to the integer) in Table 1. Calculate and annotate the deviation $\Delta = P_m - P_{ref}$ (in dB, rounded to the integer).
- Verify that:
 - o The measured power P_m decreases by at least 20 dB rotating the probe by 90°;
 - o $-6\text{dB} \leq \Delta \leq 6\text{dB}$.

Table 1: Verification of the power that the probe delivers to the receiver.

Harmonic #	Frequency MHz	P_{ref} dBm	P_m dBm	Δ dB
20	1000	-55		

7.b) If the preliminary Sample verification is successful then the Laboratory can pass to the next step, i.e. the radiated emission measurement, otherwise the Coordinator is informed and the radiated emission measurement is temporarily delayed.

7.c) Measurements performed above a reflective ground plane in the frequency range comprised between 30 and 1000 MHz (§7.3 of CISPR 16-2-3:2010) at 3 or 10 m distance.

The scope of the measurement is to obtain the best estimate and measurement uncertainty of the maximum electric field strength, in dB(μ V/m), emitted by the Sample in vertical polarization at the specified horizontal distance from the Sample (3 or 10 m) at a height between 1 and 4 m above the reflecting ground plane, see Fig. 3. The reference of the Sample for distance measurement is the plastic wall of the Sample. The use of the same measuring instrumentation as that used for radiated emission tests in the corresponding frequency range is recommended. Measurement frequencies are reported in Table 2.




Fig. 2: Measurement layout for the 30 to 1000 MHz frequency range, semi-anechoic room (3 or 10 m).

7.d) Measurements performed in a free-space environment in the frequency range comprised between 30 and 1000 MHz (§7.4 of CISPR 16-2-3:2010) and between 1000 and 6000 MHz (§7.6 of CISPR 16-2-3:2010) at 3 m distance.

The scope of the measurement is to obtain the best estimate and measurement uncertainty of the electric field strength, in dB(μ V/m), emitted by the Sample at 3 m distance from the Sample in the boresight direction. The reference of the Sample for distance measurements is the plastic wall of the Sample facing the receiving antenna. The use of the same measuring instrumentation as that used for radiated emission tests in the corresponding frequency range is recommended. Measurement frequencies are reported in Table 2.




Fig. 3: Measurement layout for the 30 MHz to 1000 MHz and for the 1000 to 6000 MHz frequency ranges in a fully-anechoic room.

7.e) It is up to the Laboratory to charge the battery before preliminary verification and measurement. Handle the Sample with care.

7.f) The EMI receiver's detector shall be set to average.

7.g) The measurement result provided by the Laboratory shall be:

- The estimate x , expressed in dB(μ V/m), of the amplitude of the selected harmonics;
- The expanded uncertainty of the estimate x , U_{lab} , expressed in dB and obtained multiplying the combined standard uncertainty by the coverage factor $k = 2$ (which corresponds to a coverage probability of about 95 % assuming normal distribution).

7.h) The Laboratory may assign a different value of U_{lab} to each measured frequency.

8. Recording electromagnetic field measurement results

8.a) The measured disturbance voltage x , in dB(μ V/m), shall be rounded up to 1 decimal figure (e.g. 68,5 dB(μ V/m)). Measurement uncertainty U_{lab} , in dB, shall be rounded up to 2 significant figures (e.g. 3,2 dB).

8.b) The values of x and U_{lab} shall be recorded in the eighth and ninth column of Table 2, respectively. The Coordinator will complete the rest of the Table 2.

Table 2: Table to be used for recording the electromagnetic field measurement result x and its measurement uncertainty U_{lab} . Columns eight and nine shall be filled up by the Laboratory, the other columns will be filled up by the Coordinator.

1	2	3	4	5	6	7	8	9	10
Band	Harmonic #	Frequency MHz	X dB(μ V/m)	U dB	x^* dB(μ V/m)	s^* dB	x dB(μ V/m)	U_{lab} dB	ζ
C	2	100	-	-	-	-			-
C	4	200	-	-	-	-			-
C	6	300	-	-	-	-			-
D	12	600	-	-	-	-			-
D	16	800	-	-	-	-			-
E	22	1100	-	-	-	-			-
E	44	2200	-	-	-	-			-
E	70	3500	-	-	-	-			-
E	104	5200	-	-	-	-			-
E	116	5800	-	-	-	-			-

8.c) The Laboratory fills columns eight and nine and sends a copy of Table 2 to the Coordinator. The Coordinator completes the rest of Table 2 and sends a copy to the Laboratory. The proficiency test result does not give evidence of any anomaly if, at all frequencies, $-2 \leq \zeta \leq 2$. Otherwise anomalies shall be described in terms of warning and action signals as discussed in §5.

NOTE: Warning signals do not add up to give an action signal.

9. Test reports

9.a) The test report issued by the Laboratory to the Coordinator shall conform to Annex D and it shall be signed by the Technical Responsible or his/her Deputy (see §-). The test report issued by the Coordinator to the Laboratory will conform to Annex E. Annexes D and E, once completed by the Laboratory and by the Coordinator, will be integral part of the present document and they will provide evidence to any interested part (e.g. the Accreditation Body) of the participation of the Laboratory to the proficiency test.

10. Remarks and complaints

10.a) The Coordinator issued and made freely available this document in order to prevent remarks and complaints from the Laboratories during the progress of the proficiency test.

10.b) Remarks and complaints will be considered by the Coordinator only if they are related to management or technical aspects actually relevant to the proficiency test but not considered in the present document. Subscription of the contract in Annex A (Italian) or B (English) implies formal acceptance of the terms and conditions of participation in the proficiency test described in this document.

10.c) Laboratories are allowed to verbally contact (e.g. by phone) the Coordinator to represent possible remarks and complaints about management and technical problems related to the proficiency test that appear during the progress of the proficiency test itself. If possible, and depending on the importance of the problem originating the remark or complaint, the Coordinator will give advice to the Laboratories in order to resolve the problem.

10.d) If the Coordinator judges that the problem cannot be verbally solved through an advice to the Laboratory then he will ask the Laboratory a written communication of the remarks and complaints. The Coordinator will discuss the remarks and complaints with his technical and scientific collaborators (see §2) and collectively take a decision about their management.

10.e) Possible technical problems related to the management of the Sample (including shipment), delay in the progress of the proficiency test caused by a Laboratory or by the Coordinator himself, can be solved by the Coordinator without involving the scientific and technical collaborators.

11. Confidentiality and impartiality

11.a) The Coordinator and his technical and scientific collaborators shall keep confidential any information pertaining the performance of the Laboratories involved in the proficiency test during its progress and after its completion. The Coordinator warrants that the results originated

from the participation of the Laboratories in the proficiency test shall be kept confidential through:

- Keeping anonymous the result associated with each Laboratory. The individual result produced by each Laboratory may be released only in such a way that the anonymity of the Laboratory is preserved.
- Keeping anonymous aggregate results (i.e., statistical average, dispersion, ...). The aggregate proficiency test results may be released only in such a way that the anonymity of the Laboratories that generated the results is preserved.
- Informing accredited Laboratories about a possible request of the Accreditation Body to reveal their proficiency test result. The proficiency test result shall be revealed to the Accreditation Body under written permission of the accredited test Laboratory.

11.b) The Coordinator and his scientific and technical collaborators will avoid any conduct that could cause some Laboratories to take advantage with respect to the others in the successful participation in the proficiency test.

11.c) Laboratories shall avoid to raise issues that could generate a situation of disparity in the successful completion in the proficiency test.

Compilare, firmare, stampare, scansionare e spedire per posta certificata a dinfo@pec.unifi.it.

CONTRATTO TRA LA Fare clic qui per immettere testo. E IL DIPARTIMENTO DI INGEGNERIA DELL'INFORMAZIONE DELL'UNIVERSITA' DI FIRENZE PER L'EFFETTUAZIONE DELLA SEGUENTE PRESTAZIONE:

“Prova valutativa di misure di emissione radiata nell’intervallo di frequenza 30-6000 MHz”

(Art. 3/C del Regolamento per lo svolgimento di attività di ricerca o didattica commissionate da terzi emanato con Decreto n. 605 del 05 giugno 2013 prot. 41287 e art. 44 del Regolamento di Amministrazione, Finanza e Contabilità emanato con Decreto n. 35026 (405) del 08 maggio 2014)

Tra

Fare clic qui per immettere testo. c.f. Fare clic qui per immettere testo. P.I. Fare clic qui per immettere testo. con sede in Fare clic qui per immettere testo.

in seguito indicata/o “Committente”, rappresentato da Fare clic qui per immettere testo.

e

il Dipartimento di Ingegneria dell'Informazione dell'Università di Firenze, c.f. e P.I. 01279680480, in seguito indicato “Unità Amministrativa”, rappresentato dal Prof. Enrico Del Re in qualità di Direttore

PREMESSA

DINFO ha le capacità per fornire servizi di circuito interlaboratorio ai Laboratori operanti nel settore delle prove di Compatibilità Elettromagnetica e il Committente intende dare evidenza della propria competenza tecnica attraverso la partecipazione a circuiti interlaboratorio.

SI CONVIENE E SI STIPULA QUANTO SEGUE

Art. 1

Oggetto del contratto

L'Unità Amministrativa effettuerà la seguente prestazione:

“Prova valutativa di misure di emissioni radiate nell’intervallo di frequenza 30-6000 MHz”, voce di tariffa “Partecipazione a circuito interlaboratorio tipo Scegliere un elemento.” (*selezionare “C” per Opzione I oppure “G” per Opzione II, vedi §3 dello schema*) approvata dal Consiglio dell'Unità Amministrativa in data 21 Marzo 2013.

Art. 2

Responsabile dell'attività

Responsabile dello svolgimento della prestazione è il Prof. Carlo Carobbi. La prestazione sarà eseguita nei locali dell'Unità Amministrativa con le modalità definite nel documento tecnico dal titolo "Proficiency test through interlaboratory comparison – Radiated emission measurements in the 30 MHz to 6000 MHz frequency range – Scheme of the proficiency test PTC(RE-30-6000-II)" (in seguito Allegato Tecnico) che accompagna il presente atto e ne costituisce parte integrante.

Art. 3

Pagamenti

Per la realizzazione della prestazione il Committente corrisponderà all'Università la somma di € Scegliere un elemento. (*selezionare 1000 per Opzione I oppure 2000 per Opzione II*) oltre IVA.

Il pagamento verrà effettuato dal Committente dietro presentazione di note di addebito, cui seguiranno regolari fatture, mediante versamenti sul:

- codice IBAN IT88A0200802837000041126939, BIC/SWIFT:UNCRITM1F86 (per pagamenti da enti privati),

oppure

- contabilità speciale 36739 (*per pagamenti da altre amm.ni in regime di tesoreria unica*) presso la Banca Unicredit S.p.A. Via Vecchietti 11 Firenze, a favore dell'Università di Firenze, Dipartimento di Ingegneria dell'Informazione, Cod. U.A. 58507, con le seguenti modalità:

- in un'unica soluzione alla stipula del presente atto.

Le note di addebito verranno inoltrate tramite pec all'indirizzo: [Fare clic qui per immettere testo.](#)

Ai fini della tracciabilità finanziaria si indica altresì che il Codice Identificativo di Gara (CIG), attribuito al presente contratto dall'Autorità di Vigilanza sui Contratti Pubblici di lavori, servizi e forniture (AVCP) su richiesta della stazione appaltante è il seguente: [Fare clic qui per immettere testo.](#) (*Tale obbligo è a carico del Committente nel caso in cui sia UN ENTE PUBBLICO. In tutti gli altri casi, si può omettere.*)

Art. 4

Risultati e proprietà intellettuale

Le parti concordano che, in base alla natura del servizio, non si prevede che possano derivare da questa attività invenzioni brevettabili.

Gli eventuali risultati della ricerca condivisa saranno pubblicati dopo che tutte le parti si saranno accordate sui termini e le condizioni della pubblicazione medesima.

Art. 5

Durata e termini di esecuzione del servizio

La prestazione avrà inizio dalla data di stipula del presente accordo e verrà effettuata entro un anno secondo la procedura descritta nell'Allegato Tecnico, che costituisce parte integrante del presente Contratto. Firmando questo Contratto il Committente accetta i termini di svolgimento del servizio descritti nell'Allegato Tecnico.

L'importo delle prestazioni e la durata può essere estesa attraverso un nuovo accordo sottoscritto dalle parti.

Art. 6

Riservatezza e pubblicità

Il Dipartimento ed il personale coinvolto sono tenuti a rispettare gli obblighi di non concorrenza e riservatezza (le informazioni che devono essere considerate riservate sono specificate nell'Allegato Tecnico).

Art. 7

Disposizioni finali e Foro Competente

Per tutto quanto non espressamente stabilito, restano ferme le disposizioni previste dal Codice Civile. Tutte le eventuali dispute connesse all'esecuzione del presente contratto dovranno essere risolte in via amichevole fra le parti. In caso ciò non risultasse possibile, si dichiara sin d'ora che deve considerarsi foro esclusivamente competente il Tribunale di Firenze

Art. 8

Spese del contratto

Il presente atto verrà registrato solo in caso d'uso ai sensi dell'art. 5, II comma, del D.P.R. n. 131 del 26/4/1986 e successive modifiche, a cura e spese della parte richiedente.

Le spese di bollo sono a carico del Committente

p. il COMMITTENTE Fare clic qui per immettere testo., lì Fare clic qui per immettere testo.
(Fare clic qui per immettere testo.)

..... (firma)

p. l'UNITA' AMMINISTRATIVA Firenze, lì Fare clic qui per immettere testo.
(Il Direttore Prof. Enrico Del Re)

..... (firma)

Per presa visione,

il responsabile dell'attività

(Prof. Carlo Carobbi)

..... (firma)

Per presa visione ed espressa e separata accettazione, anche ai sensi e per gli effetti dell'art. 1341 c.c., delle clausole sub. 3 (pagamenti), 4 (risultati e proprietà intellettuale), 5 (durata e termini di esecuzione del servizio), 6 (riservatezza e pubblicità), 7 (disposizioni finali e Foro competente).

p. il COMMITTENTE Fare clic qui per immettere testo., lì Fare clic qui per immettere testo.

(Fare clic qui per immettere testo.)

..... (firma)

Annex B

Fill, sign, print, scan and send by certified e-mail to dinfo@pec.unifi.it.

CONTRACT BETWEEN Fare clic qui per immettere testo. AND THE DEPARTMENT OF INFORMATION ENGINEERING OF THE UNIVERSITY OF FLORENCE FOR THE FOLLOWING SERVICE

“Proficiency test of radiated emission measurements in the 30-6000 MHz frequency range”

Between

Fare clic qui per immettere testo., tax identification number Fare clic qui per immettere testo., with premises in Fare clic qui per immettere testo. hereinafter referred to as “the Laboratory”, represented by Fare clic qui per immettere testo.

and

The Department of Information Engineering of the University of Florence, fiscal code and VAT number 01279680480, hereinafter referred to as “DINFO”, represented by Prof. Enrico Del Re, in his capacity as head of DINFO

whereas

DINFO has the capability to provide the interlaboratory comparison service to Laboratories operating in the sector of Electromagnetic Compatibility testing and the Laboratory is willing to give evidence of his technical competence through participation to interlaboratory comparisons.

the following agreement is drawn-up

Art.1. – Subject of the Contract

DINFO will carry out the following service “Proficiency test of conducted emission measurements in the 9 kHz to 30 MHz frequency range” price list item “Participaton in interlaboratory comparison type Scegliere un elemento.” (*select “C” for Option I or “G” for Option II, see §3 of the scheme*) as approved by the Board of DINFO in March 21, 2013.

Art. 2. Responsibility of the service

The person (Responsible) in charge of carrying out the service is Prof. Carlo Carobbi. The service will be carried out in the DINFO premises according to the procedure defined in the technical document titled “Proficiency test through interlaboratory comparison – Radiated emission measurements in the 30-6000 MHz frequency range – Scheme of the proficiency test PTC(RE-30-6000-II),” (for brevity “Technical Annex” in the following) which is an integral part of this Contract.

Art. 3. Fees

In order to obtain the service specified in the Technical Annex, the Laboratory will pay the sum of € Scegliere un elemento. *(insert 1000 if Option I is selected, or 2000 if Option II is selected)* plus VAT.

The sum will be paid to DINFO under the following terms of payment:

- The whole amount upon signing this Contract.

All payments will be made by the Laboratory, on presentation of debit notes followed by regular invoices, addressed to:

Bank name: UNICREDIT Banca

Bank's address: Via Vecchietti 11 – Firenze

Account holder: University of Florence - Department of Information Engineering - (cod. UA. 58507)

Account Number: 41126939

IBAN: IT88A0200802837000041126939

BIC/SWIFT: UNCRITM1F86

Art. 4. Results and Intellectual property

The parties agree that, due to the nature of the service, it is not expected that patentable inventions can arise from this activity.

Possible joint results of the research will be published after both parties have agreed about the publication terms.

Art. 5. Duration and terms of execution of the service

The service will be completed within one year starting from the date of drawing up of this Contract. The service will be performed according to the procedure described in the Technical Annex which is an integral part of this Contract. By signing this Contract the Laboratory agrees on the terms of execution of the service as described in the Technical Annex. The amount of the services and the duration can be extended through an agreement signed by the parties.

Art. 6. Confidentiality and publicity

DINFO, the Laboratory and the staff involved are bound to respect the obligations of non rivalry and confidentiality (possible details about which information must be considered confidential are specified in the Technical Annex).

Art. 7. Final dispositions

For whatsoever has not been expressly agreed, the Contract shall be governed by the law of the country where the party required to effect the characteristic performance has his habitual residence. Where it is impossible to determine the characteristic performance of the contract, it shall have regard to the law of the country with which it is most closely connected. All disputes or differences between the Parties arising out or in connection with this Agreement which the Parties cannot settle amicably shall be finally submitted to the jurisdiction of the defendant, that is *Fare clic qui per immettere testo*. if the Laboratory is the defendant, Florence Court if the University of Florence is the defendant.

Art. 8. Cost of the Contract

This Contract will be registered only in the case of use according to art. 5, paragraph II of the D.P.R. 26/4/1986 n. 131 and subsequent modifications. The Laboratory is responsible for the necessary arrangements and expenses, including the cost of stamps.

-----00----

For the Laboratory

Fare clic qui per immettere testo., Fare clic qui per immettere testo.

(Fare clic qui per immettere testo.

.....(*signature*)

For the Department of Information Engineering

Florence, Fare clic qui per immettere testo.

(Prof. Enrico Del Re)

.....(*signature*)

Signature of acknowledgment of the Responsible of the service

Prof. Carlo Carobbi

.....(*signature*)

Annex C

**Proficiency test through interlaboratory comparison –
Radiated emission measurements in the 30-6000 MHz frequency range**

Name of the Laboratory (mandatory):
.....

Technical Responsible (mandatory)

First name: Last name:

E-mail:

Phone:

Cell phone (optional):

Deputy of the Technical Responsible (optional):

First name: Last name:

E-mail:

Phone:

Cell phone (optional):

Shipping address (mandatory):

Address:
.....

ZIP Code:

City:

Country:

Date:/...../.....

Test report issued by the participating Laboratory

Laboratory: Name of the Laboratory

Laboratory Code: PTC(RE-30-6000-II)LAB(#)

Address: Address of the Laboratory

Technical Responsible: First name and last name of the Technical Responsible or his/her Deputy

E-mail: E-mail address of the Technical Responsible or his/her Deputy

Phone: Phone number of the Technical Responsible or his/her Deputy

Date of issue: Date of issue of this test report

Date of Sample receipt:.....

Date of measurements:

Data of Sample shipment:

Test result

Fill in the empty cells of columns eight and nine with the measured value x and the measurement uncertainty U_{lab} .

1	2	3	4	5	6	7	8	9	10
Band	Harmonic #	Frequency MHz	X dB(μ V/m)	U dB	x^* dB(μ V/m)	s^* dB	x dB(μ V/m)	U_{lab} dB	ζ
C	2	100	-	-	-	-			-
C	4	200	-	-	-	-			-
C	6	300	-	-	-	-			-
D	12	600	-	-	-	-			-
D	16	800	-	-	-	-			-
E	22	1100	-	-	-	-			-
E	44	2200	-	-	-	-			-
E	70	3500	-	-	-	-			-
E	104	5200	-	-	-	-			-
E	116	5800	-	-	-	-			-

Comments: Comments may be inserted here (optional)

Photos: At least one photo shall be inserted here (mandatory).

Sign of the Technical Responsible or his/her Deputy

.....

Test report no. XYZ
Issued by the Coordinator of the proficiency test code PTC(RE-30-6000-II)

Carlo Carobbi
 Dipartimento di Ingegneria dell'Informazione
 Università degli Studi di Firenze
 Via S. Marta, 3 – 50139 Firenze
 Phone: +39 055 2758501
 Mob. phone: +39 329 6509116
 e-mail: carlo.carobbi@unifi.it

to the participating Laboratory

Laboratory: Name of the Laboratory
 Laboratory Code: PTC(RE-30-6000-II)LAB(#)
 Address: Address of the Laboratory

Start and stop dates of the proficiency test:
 Number of participants:
 Date of measurements of the participating Laboratory:
 Date of issue of this report:

Test result

The cells of columns eight and nine are filled in by the Laboratory, the other ones are filled by the Coordinator.

1	2	3	4	5	6	7	8	9	10
Band	Harmonic #	Frequency MHz	X dB(μ V/m)	U dB	x^* dB(μ V/m)	s^* dB	x dB(μ V/m)	U_{lab} dB	ζ
C	2	100	-	-	-	-			-
C	4	200	-	-	-	-			-
C	6	300	-	-	-	-			-
D	12	600	-	-	-	-			-
D	16	800	-	-	-	-			-
E	22	1100	-	-	-	-			-
E	44	2200	-	-	-	-			-
E	70	3500	-	-	-	-			-
E	104	5200	-	-	-	-			-
E	116	5800	-	-	-	-			-

Outcome

Here the Coordinator inserts one of the following outcomes:

- No anomaly is detected
- Warning signal(s) is (are) detected
- Action signal(s) is (are) detected

Sign of the Coordinator

.....